



October 17, 2025

Glen Johnson
FEH BOCES_Salmon River Central School
District
637 County Route 1
Fort Covington, NY 12937

RE: Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Dear Glen Johnson:

Enclosed are the analytical results for sample(s) received by the laboratory on October 03, 2025. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Michelle Cohen
michelle.cohen@pacelabs.com
516-370-6000
Project Manager

Enclosures

cc: Ray Filley, FEH BOCES_Salmon River Central School
District
Tiffany O'Connor, FEH BOCES_Salmon River Central
School District
Linda Shaw, FEH BOCES_Salmon River Central School
District
Ashley White, FEH BOCES_Salmon River Central School
District



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-01		Lab ID: 70383633001		Collected: 09/30/25 07:22		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:10	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-02		Lab ID: 70383633002		Collected: 09/30/25 07:21		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:12	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-03		Lab ID: 70383633003		Collected: 09/30/25 07:23		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:13	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-04		Lab ID: 70383633004		Collected: 09/30/25 07:24		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.0	ug/L	1.0	1		10/15/25 15:15	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-05		Lab ID: 70383633005		Collected: 09/30/25 07:28		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:16	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-06		Lab ID: 70383633006		Collected: 09/30/25 07:27		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:21	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-07		Lab ID: 70383633007		Collected: 09/30/25 07:42		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:22	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-08		Lab ID: 70383633008		Collected: 09/30/25 07:31		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:24	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-09		Lab ID: 70383633009		Collected: 09/30/25 07:32		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:25	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-10		Lab ID: 70383633010		Collected: 09/30/25 07:32		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:27	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-11		Lab ID: 70383633011		Collected: 09/30/25 07:34		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:28	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-13		Lab ID: 70383633012		Collected: 09/30/25 07:36		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:30	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-15		Lab ID: 70383633013		Collected: 09/30/25 07:37		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:31	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-16		Lab ID: 70383633014		Collected: 09/30/25 07:16		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:33	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-17		Lab ID: 70383633015		Collected: 09/30/25 07:15		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:34	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-18		Lab ID: 70383633016		Collected: 09/30/25 07:15		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:38	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-19		Lab ID: 70383633017		Collected: 09/30/25 07:14		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:40	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-20		Lab ID: 70383633018		Collected: 09/30/25 07:12		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:44	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-21		Lab ID: 70383633019		Collected: 09/30/25 07:13		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:49	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-22		Lab ID: 70383633020		Collected: 09/30/25 07:04		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:56	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-23		Lab ID: 70383633021		Collected: 09/30/25 07:05		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 15:57	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-24		Lab ID: 70383633022		Collected: 09/30/25 07:10		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		10/15/25 15:59	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-25		Lab ID: 70383633023		Collected: 09/30/25 06:47		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:00	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-26		Lab ID: 70383633024		Collected: 09/30/25 06:47		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:02	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-27		Lab ID: 70383633025		Collected: 09/30/25 06:50		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.6	ug/L	1.0	1		10/15/25 16:03	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-28		Lab ID: 70383633026		Collected: 09/30/25 06:50		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.4	ug/L	1.0	1		10/15/25 16:05	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-29		Lab ID: 70383633027		Collected: 09/30/25 06:51		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		10/15/25 16:06	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-30		Lab ID: 70383633028		Collected: 09/30/25 06:51		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		10/15/25 16:07	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-31		Lab ID: 70383633029		Collected: 09/30/25 06:54		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:09	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-32		Lab ID: 70383633030		Collected: 09/30/25 06:54		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.9	ug/L	1.0	1		10/15/25 16:13	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-33		Lab ID: 70383633031		Collected: 09/30/25 06:56		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	2.2	ug/L	1.0	1		10/15/25 16:15	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-35		Lab ID: 70383633032		Collected: 09/30/25 06:57		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:16	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-36		Lab ID: 70383633033		Collected: 09/30/25 06:57		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.9	ug/L	1.0	1		10/15/25 16:18	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-37		Lab ID: 70383633034		Collected: 09/30/25 07:00		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		10/15/25 16:19	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-38		Lab ID: 70383633035		Collected: 09/30/25 07:00		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:21	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-39		Lab ID: 70383633036		Collected: 09/30/25 07:01		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	2.4	ug/L	1.0	1		10/15/25 16:22	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-41		Lab ID: 70383633037		Collected: 09/30/25 06:45		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	2.1	ug/L	1.0	1		10/15/25 16:24	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-43		Lab ID: 70383633038		Collected: 09/30/25 06:44		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		10/15/25 16:33	7439-92-1	M1	

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-44		Lab ID: 70383633039		Collected: 09/30/25 06:41		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:37	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-45		Lab ID: 70383633040		Collected: 09/30/25 06:42		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:41	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-46		Lab ID: 70383633041		Collected: 09/30/25 06:39		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		10/15/25 16:42	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-48		Lab ID: 70383633042		Collected: 09/30/25 06:29		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:44	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-49		Lab ID: 70383633043		Collected: 09/30/25 06:30		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:48	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-50		Lab ID: 70383633044		Collected: 09/30/25 06:27		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:50	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-51		Lab ID: 70383633045		Collected: 09/30/25 06:28		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:51	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-52		Lab ID: 70383633046		Collected: 09/30/25 06:26		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		10/15/25 16:53	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-53		Lab ID: 70383633047		Collected: 09/30/25 06:25		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:54	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-54		Lab ID: 70383633048		Collected: 09/30/25 06:24		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:56	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-55		Lab ID: 70383633049		Collected: 09/30/25 06:23		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
		Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:57	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-56		Lab ID: 70383633050		Collected: 09/30/25 06:22		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 16:59	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-57		Lab ID: 70383633051		Collected: 09/30/25 06:19		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:00	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-58		Lab ID: 70383633052		Collected: 09/30/25 06:20		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:02	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-59		Lab ID: 70383633053		Collected: 09/30/25 06:21		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:06	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-62		Lab ID: 70383633054		Collected: 09/30/25 06:14		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:07	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-63		Lab ID: 70383633055		Collected: 09/30/25 06:15		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:09	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-65		Lab ID: 70383633056		Collected: 09/30/25 06:07		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.2	ug/L	1.0	1		10/15/25 17:10	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-66		Lab ID: 70383633057		Collected: 09/30/25 06:09		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:12	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-67		Lab ID: 70383633058		Collected: 09/30/25 06:05		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:16	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-68		Lab ID: 70383633059		Collected: 09/30/25 07:59		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:24	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-69		Lab ID: 70383633060		Collected: 09/30/25 07:54		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:28	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-70		Lab ID: 70383633061		Collected: 09/30/25 07:55		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:29	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-71		Lab ID: 70383633062		Collected: 09/30/25 07:56		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:31	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-72		Lab ID: 70383633063		Collected: 09/30/25 07:45		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.2	ug/L	1.0	1		10/15/25 17:32	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-73		Lab ID: 70383633064		Collected: 09/30/25 07:45		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:34	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-75		Lab ID: 70383633065		Collected: 09/30/25 07:50		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:35	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-76		Lab ID: 70383633066		Collected: 09/30/25 06:34		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:37	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-77		Lab ID: 70383633067		Collected: 09/30/25 06:37		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:41	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-107		Lab ID: 70383633068		Collected: 09/30/25 06:36		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:43	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-74		Lab ID: 70383633069		Collected: 10/02/25 05:59		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		10/15/25 17:44	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-78		Lab ID: 70383633070		Collected: 10/02/25 06:00		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:46	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-79		Lab ID: 70383633071		Collected: 10/02/25 06:01		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:47	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-80		Lab ID: 70383633072		Collected: 10/02/25 06:04		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:48	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-81		Lab ID: 70383633073		Collected: 10/02/25 06:04		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:50	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-82		Lab ID: 70383633074		Collected: 10/02/25 06:08		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/15/25 17:51	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-83		Lab ID: 70383633075		Collected: 10/02/25 06:08		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:06	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-84		Lab ID: 70383633076		Collected: 10/02/25 06:40		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:15	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-85		Lab ID: 70383633077		Collected: 10/02/25 06:40		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:19	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-86		Lab ID: 70383633078		Collected: 10/02/25 06:41		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:21	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-87		Lab ID: 70383633079		Collected: 10/02/25 06:41		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:22	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-88		Lab ID: 70383633080		Collected: 10/02/25 06:25		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:24	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-89		Lab ID: 70383633081		Collected: 10/02/25 06:25		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:25	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-90		Lab ID: 70383633082		Collected: 10/02/25 06:27		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:27	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-91		Lab ID: 70383633083		Collected: 10/02/25 06:21		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:28	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-94		Lab ID: 70383633084		Collected: 10/02/25 06:16		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:33	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-95		Lab ID: 70383633085		Collected: 10/02/25 06:14		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:34	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-96		Lab ID: 70383633086		Collected: 10/02/25 06:14		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:35	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-99		Lab ID: 70383633087		Collected: 10/02/25 06:31		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:37	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-100		Lab ID: 70383633088		Collected: 10/02/25 06:31		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:38	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-103		Lab ID: 70383633089		Collected: 10/02/25 06:46		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:40	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-104		Lab ID: 70383633090		Collected: 10/02/25 06:45		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:54	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-105		Lab ID: 70383633091		Collected: 10/02/25 06:49		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 12:59	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-110		Lab ID: 70383633092		Collected: 10/02/25 05:59		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:03	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-111		Lab ID: 70383633093		Collected: 10/02/25 06:06		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:08	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-112		Lab ID: 70383633094		Collected: 10/02/25 06:04		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:09	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-113		Lab ID: 70383633095		Collected: 10/02/25 06:10		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:11	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-114		Lab ID: 70383633096		Collected: 10/02/25 06:10		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:12	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-115		Lab ID: 70383633097		Collected: 10/02/25 06:10		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:13	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-116		Lab ID: 70383633098		Collected: 10/02/25 06:24		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:15	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30
Pace Project No.: 70383633

Sample: SRC-117		Lab ID: 70383633099		Collected: 10/02/25 06:24		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.3	ug/L	1.0	1		10/16/25 13:16	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-118		Lab ID: 70383633100		Collected: 10/02/25 06:25		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:18	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-119		Lab ID: 70383633101		Collected: 10/02/25 06:25		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:19	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-120		Lab ID: 70383633102		Collected: 10/02/25 06:28		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.0	ug/L	1.0	1		10/16/25 13:21	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-121		Lab ID: 70383633103		Collected: 10/02/25 06:33		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:25	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Sample: SRC-122		Lab ID: 70383633104		Collected: 10/02/25 06:46		Received: 10/03/25 06:00		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		10/16/25 13:27	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

QC Batch:	423505	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70383633001, 70383633002, 70383633003, 70383633004, 70383633005, 70383633006, 70383633007, 70383633008, 70383633009, 70383633010, 70383633011, 70383633012, 70383633013, 70383633014, 70383633015, 70383633016, 70383633017		

METHOD BLANK:	2257152	Matrix:	Water
Associated Lab Samples:	70383633001, 70383633002, 70383633003, 70383633004, 70383633005, 70383633006, 70383633007, 70383633008, 70383633009, 70383633010, 70383633011, 70383633012, 70383633013, 70383633014, 70383633015, 70383633016, 70383633017		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/15/25 14:54	

LABORATORY CONTROL SAMPLE:	2257153					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.8	104	85-115	

MATRIX SPIKE SAMPLE:		2257155					
		70383632059	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	60.3	120	70-130	

MATRIX SPIKE SAMPLE:		2257157					
		70383632060	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	53.5	107	70-130	

SAMPLE DUPLICATE:	2257154				
Parameter	Units	70383632059 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2257156				
Parameter	Units	70383632060 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

QC Batch:	423511	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70383633018, 70383633019, 70383633020, 70383633021, 70383633022, 70383633023, 70383633024, 70383633025, 70383633026, 70383633027, 70383633028, 70383633029, 70383633030, 70383633031, 70383633032, 70383633033, 70383633034, 70383633035, 70383633036, 70383633037		

METHOD BLANK:	2257190	Matrix:	Water
Associated Lab Samples:	70383633018, 70383633019, 70383633020, 70383633021, 70383633022, 70383633023, 70383633024, 70383633025, 70383633026, 70383633027, 70383633028, 70383633029, 70383633030, 70383633031, 70383633032, 70383633033, 70383633034, 70383633035, 70383633036, 70383633037		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/15/25 15:41	

LABORATORY CONTROL SAMPLE:	2257191					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	46.6	93	85-115	

MATRIX SPIKE SAMPLE:		2257193					
		70383633018	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	55.1	110	70-130	

MATRIX SPIKE SAMPLE:		2257195					
		70383633019	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	64.6	129	70-130	

SAMPLE DUPLICATE:	2257192				
Parameter	Units	70383633018 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2257194				
Parameter	Units	70383633019 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

QC Batch:	423535	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70383633038, 70383633039, 70383633040, 70383633041, 70383633042, 70383633043, 70383633044, 70383633045, 70383633046, 70383633047, 70383633048, 70383633049, 70383633050, 70383633051, 70383633052, 70383633053, 70383633054, 70383633055, 70383633056, 70383633057		

METHOD BLANK: 2257316

Matrix: Water

Associated Lab Samples: 70383633038, 70383633039, 70383633040, 70383633041, 70383633042, 70383633043, 70383633044, 70383633045, 70383633046, 70383633047, 70383633048, 70383633049, 70383633050, 70383633051, 70383633052, 70383633053, 70383633054, 70383633055, 70383633056, 70383633057

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/15/25 16:27	

LABORATORY CONTROL SAMPLE: 2257317

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.5	101	85-115	

MATRIX SPIKE SAMPLE: 2257319

Parameter	Units	70383633038 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.7	50	67.3	131	70-130	M1

MATRIX SPIKE SAMPLE: 2257321

Parameter	Units	70383633039 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	62.0	123	70-130	

SAMPLE DUPLICATE: 2257318

Parameter	Units	70383633038 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	1.7	1.6	1	

SAMPLE DUPLICATE: 2257320

Parameter	Units	70383633039 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

QC Batch:	423537	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70383633058, 70383633059, 70383633060, 70383633061, 70383633062, 70383633063, 70383633064, 70383633065, 70383633066, 70383633067, 70383633068, 70383633069, 70383633070, 70383633071, 70383633072, 70383633073, 70383633074		

METHOD BLANK:	2257322	Matrix:	Water
Associated Lab Samples:	70383633058, 70383633059, 70383633060, 70383633061, 70383633062, 70383633063, 70383633064, 70383633065, 70383633066, 70383633067, 70383633068, 70383633069, 70383633070, 70383633071, 70383633072, 70383633073, 70383633074		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/15/25 17:13	

LABORATORY CONTROL SAMPLE:	2257323					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	50.4	101	85-115	

MATRIX SPIKE SAMPLE:	2257325						
Parameter	Units	70383633058 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	58.7	117	70-130	

MATRIX SPIKE SAMPLE:	2257327						
Parameter	Units	70383633059 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	63.6	127	70-130	

SAMPLE DUPLICATE: 2257324

Parameter	Units	70383633058 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 2257326

Parameter	Units	70383633059 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

QC Batch:	423677	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70383633075, 70383633076, 70383633077, 70383633078, 70383633079, 70383633080, 70383633081, 70383633082, 70383633083, 70383633084, 70383633085, 70383633086, 70383633087, 70383633088, 70383633089		

METHOD BLANK:	2258178	Matrix:	Water
Associated Lab Samples:	70383633075, 70383633076, 70383633077, 70383633078, 70383633079, 70383633080, 70383633081, 70383633082, 70383633083, 70383633084, 70383633085, 70383633086, 70383633087, 70383633088, 70383633089		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/16/25 12:03	

LABORATORY CONTROL SAMPLE:	2258179					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	47.2	94	85-115	

MATRIX SPIKE SAMPLE:	2258182						
Parameter	Units	70383633075 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	47.6	95	70-130	

MATRIX SPIKE SAMPLE:	2258184						
Parameter	Units	70383633076 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	48.3	96	70-130	

SAMPLE DUPLICATE:	2258181				
Parameter	Units	70383633075 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2258183				
Parameter	Units	70383633076 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

QC Batch:	423678	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70383633090, 70383633091, 70383633092, 70383633093, 70383633094, 70383633095, 70383633096, 70383633097, 70383633098, 70383633099, 70383633100, 70383633101, 70383633102, 70383633103, 70383633104		

METHOD BLANK:	2258186	Matrix:	Water
Associated Lab Samples:	70383633090, 70383633091, 70383633092, 70383633093, 70383633094, 70383633095, 70383633096, 70383633097, 70383633098, 70383633099, 70383633100, 70383633101, 70383633102, 70383633103, 70383633104		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	10/16/25 12:52	

LABORATORY CONTROL SAMPLE:	2258187					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	47.6	95	85-115	

MATRIX SPIKE SAMPLE:		2258189					
		70383633090	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	52.0	104	70-130	

MATRIX SPIKE SAMPLE:		2258191					
		70383633091	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	<1.0	50	48.4	96	70-130	

SAMPLE DUPLICATE:	2258188				
Parameter	Units	70383633090 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE:	2258190				
Parameter	Units	70383633091 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

ANALYTE QUALIFIERS

M1 Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70383633001	SRC-01	EPA 200.8	423505		
70383633002	SRC-02	EPA 200.8	423505		
70383633003	SRC-03	EPA 200.8	423505		
70383633004	SRC-04	EPA 200.8	423505		
70383633005	SRC-05	EPA 200.8	423505		
70383633006	SRC-06	EPA 200.8	423505		
70383633007	SRC-07	EPA 200.8	423505		
70383633008	SRC-08	EPA 200.8	423505		
70383633009	SRC-09	EPA 200.8	423505		
70383633010	SRC-10	EPA 200.8	423505		
70383633011	SRC-11	EPA 200.8	423505		
70383633012	SRC-13	EPA 200.8	423505		
70383633013	SRC-15	EPA 200.8	423505		
70383633014	SRC-16	EPA 200.8	423505		
70383633015	SRC-17	EPA 200.8	423505		
70383633016	SRC-18	EPA 200.8	423505		
70383633017	SRC-19	EPA 200.8	423505		
70383633018	SRC-20	EPA 200.8	423511		
70383633019	SRC-21	EPA 200.8	423511		
70383633020	SRC-22	EPA 200.8	423511		
70383633021	SRC-23	EPA 200.8	423511		
70383633022	SRC-24	EPA 200.8	423511		
70383633023	SRC-25	EPA 200.8	423511		
70383633024	SRC-26	EPA 200.8	423511		
70383633025	SRC-27	EPA 200.8	423511		
70383633026	SRC-28	EPA 200.8	423511		
70383633027	SRC-29	EPA 200.8	423511		
70383633028	SRC-30	EPA 200.8	423511		
70383633029	SRC-31	EPA 200.8	423511		
70383633030	SRC-32	EPA 200.8	423511		
70383633031	SRC-33	EPA 200.8	423511		
70383633032	SRC-35	EPA 200.8	423511		
70383633033	SRC-36	EPA 200.8	423511		
70383633034	SRC-37	EPA 200.8	423511		
70383633035	SRC-38	EPA 200.8	423511		
70383633036	SRC-39	EPA 200.8	423511		
70383633037	SRC-41	EPA 200.8	423511		
70383633038	SRC-43	EPA 200.8	423535		
70383633039	SRC-44	EPA 200.8	423535		
70383633040	SRC-45	EPA 200.8	423535		
70383633041	SRC-46	EPA 200.8	423535		
70383633042	SRC-48	EPA 200.8	423535		
70383633043	SRC-49	EPA 200.8	423535		
70383633044	SRC-50	EPA 200.8	423535		
70383633045	SRC-51	EPA 200.8	423535		
70383633046	SRC-52	EPA 200.8	423535		
70383633047	SRC-53	EPA 200.8	423535		
70383633048	SRC-54	EPA 200.8	423535		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70383633049	SRC-55	EPA 200.8	423535		
70383633050	SRC-56	EPA 200.8	423535		
70383633051	SRC-57	EPA 200.8	423535		
70383633052	SRC-58	EPA 200.8	423535		
70383633053	SRC-59	EPA 200.8	423535		
70383633054	SRC-62	EPA 200.8	423535		
70383633055	SRC-63	EPA 200.8	423535		
70383633056	SRC-65	EPA 200.8	423535		
70383633057	SRC-66	EPA 200.8	423535		
70383633058	SRC-67	EPA 200.8	423537		
70383633059	SRC-68	EPA 200.8	423537		
70383633060	SRC-69	EPA 200.8	423537		
70383633061	SRC-70	EPA 200.8	423537		
70383633062	SRC-71	EPA 200.8	423537		
70383633063	SRC-72	EPA 200.8	423537		
70383633064	SRC-73	EPA 200.8	423537		
70383633065	SRC-75	EPA 200.8	423537		
70383633066	SRC-76	EPA 200.8	423537		
70383633067	SRC-77	EPA 200.8	423537		
70383633068	SRC-107	EPA 200.8	423537		
70383633069	SRC-74	EPA 200.8	423537		
70383633070	SRC-78	EPA 200.8	423537		
70383633071	SRC-79	EPA 200.8	423537		
70383633072	SRC-80	EPA 200.8	423537		
70383633073	SRC-81	EPA 200.8	423537		
70383633074	SRC-82	EPA 200.8	423537		
70383633075	SRC-83	EPA 200.8	423677		
70383633076	SRC-84	EPA 200.8	423677		
70383633077	SRC-85	EPA 200.8	423677		
70383633078	SRC-86	EPA 200.8	423677		
70383633079	SRC-87	EPA 200.8	423677		
70383633080	SRC-88	EPA 200.8	423677		
70383633081	SRC-89	EPA 200.8	423677		
70383633082	SRC-90	EPA 200.8	423677		
70383633083	SRC-91	EPA 200.8	423677		
70383633084	SRC-94	EPA 200.8	423677		
70383633085	SRC-95	EPA 200.8	423677		
70383633086	SRC-96	EPA 200.8	423677		
70383633087	SRC-99	EPA 200.8	423677		
70383633088	SRC-100	EPA 200.8	423677		
70383633089	SRC-103	EPA 200.8	423677		
70383633090	SRC-104	EPA 200.8	423678		
70383633091	SRC-105	EPA 200.8	423678		
70383633092	SRC-110	EPA 200.8	423678		
70383633093	SRC-111	EPA 200.8	423678		
70383633094	SRC-112	EPA 200.8	423678		
70383633095	SRC-113	EPA 200.8	423678		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: SALMON RIVER CSD 9/30

Pace Project No.: 70383633

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70383633096	SRC-114	EPA 200.8	423678		
70383633097	SRC-115	EPA 200.8	423678		
70383633098	SRC-116	EPA 200.8	423678		
70383633099	SRC-117	EPA 200.8	423678		
70383633100	SRC-118	EPA 200.8	423678		
70383633101	SRC-119	EPA 200.8	423678		
70383633102	SRC-120	EPA 200.8	423678		
70383633103	SRC-121	EPA 200.8	423678		
70383633104	SRC-122	EPA 200.8	423678		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



Pace* Location Requested (City/State):
Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Company Name: FEH BOCES Salmon River CSD
Street Address: 637 Co Rd 1, Fort Covington, NY 12937

Contact/Report To: Glen Johnson
Phone #: 5672429783
E-Mail: gjohnson@srk12.org
Cc E-Mail:

Customer Project #: Salmon River Central School District

Invoice To: Ashley White
Invoice E-Mail: acwhite@srk12.org

Site Collection Info/Facility ID (as applicable):

Purchase Order # (if applicable):
Quote #:

Time Zone Collected: [] AK [] MT [] PT [] CT [X] ET

Country / State origin of sample(s): New York

Data Deliverables: Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

[] Level II [] Level III [] Level IV

[] LEQUIS [] Other: _____

Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other: _____

Date Results Requested: Standard 10 business day

DW PWSID # or WW Permit # as applicable:
Field Filtered (if applicable): [] Yes [] No
Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OI), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Composite End		Number & Type of Containers	200.8 Drinking Water (Pb only)	Preservation non-conformance identified for sample.
			Date	Time		Date	Time			
SRC-01	DW	G	9/30/25	0722				1	X	
SRC-02				0721						
SRC-03				0723						
SRC-04				0724						
SRC-05				0728						
SRC-06				0727						
SRC-07				0742						
SRC-08				0731						
SRC-09				0732						
SRC-10				0732						

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: Lauren Doherty

Printed Name: Signature:

Relinquished by/Company: (Signature) Date/Time: 10/2/25 12:00

Received by/Company: (Signature) Date/Time: 10/2/25 12:00

Relinquished by/Company: (Signature) Date/Time: 10/2 12:10

Received by/Company: (Signature) Date/Time: 10/3/25 6:00

Relinquished by/Company: (Signature) Date/Time: 10/3/25 6:00

Received by/Company: (Signature) Date/Time: 10/3/25 6:00

Relinquished by/Company: (Signature) Date/Time: 10/3/25 6:00

Received by/Company: (Signature) Date/Time: 10/3/25 6:00

Submitting a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace* Terms and Conditions found at https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/

WO#: 70383633



70383633

Specify Container Size **

Identify Container: Preservative Type***

Analysis Requested

*** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) EnCore, (8) TerraCore, (9) Other
*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other
Proj. Mgr: Alexandria Correa
ActNum / Client ID:
Table #:
Profile / Template:
Prelg / Bottle Ord. ID:
Sample Comment

Additional Instructions from Pace*:

Coolers: 12/11 Correction Factor (°C): 10.1 Obs. Temp. (°C): 14.8 Corrected Temp. (°C): 14.9

Tracking Number: 1012125 12:00

Delivered by: [] In-Person [] Courier

[] FedEx [] UPS [] Other

Page: 1 of 1

CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Company Name: FEH BOCES Salmon River CSD
Street Address: 637 Co Rd 1, Fort Covington, NY 12937

Contact/Report To: **Glen Johnson**
Phone #: **5672429783**
E-Mail: **gjohnson@srk.com**

Customer Project #:
Project Name: Salmon River Central School District

Invoice To: **Ashley White**
Invoice E-Mail: **acwhite@srk12.org**

Site Collection Info/Facility ID (as applicable):

Purchase Order # (if applicable):

Quote #:

County / State origin of sample(s):

County / State origin of sample(s):	New York
(DW, RCRA, etc.) as applicable: NY Lead in School DW	

Rush (Pre-approval required): ☐ 2 Day ☐ 3 day ☐ 5 day ☐ Other _____	DW PWSID # or WW Permit # as applicable: _____
Date Results Remitted: _____ Standard 10 business day	Field Filtered (if applicable): ☐ Yes ☐ No Analysis: _____

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Biossaw (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers
			Date	Time	Date	Time		
SRC-31	DW	G	9/30/25	0654				1
SRC-32				0654				
SRC-33				0656				
SRC-34								
SRC-35				0657				
SRC-36				0657				
SRC-37				0700				
SRC-38				0700				
SRC-39				0701				
SRC-40								

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: _____
Printed Name: Lauren Doherty
Signature: 

[illegible]

Relinquished by/Company: (Signature)

Date/Time:

Relinquished by/Company: (Signature)

Date/Time:

Relinquished by/Company: (Signature)

Date/Time:

Relinquished by/Company: (Signature)

Date/Time:

Delivered by: ☐ In-Person ☐ Courier☐ FedEx ☐ UPS ☐ Other

Page: 4 of 11

CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Login Label Here



Scan QR Code for instructions

Company Name: FEH BOCES Salmon River CSD
Street Address: 637 Co Rd 1, Fort Covington, NY 12937
Customer Project #: Salmon River Central School District
Project Name: Salmon River Central School District
Site Collection Info/Facility ID (as applicable):
Contact/Report To: Glen Johnson
Phone #: 5672429783
E-Mail: gjohnson@srk12.org
Cc E-Mail:
Invoice To: Ashley White
Invoice E-Mail: acwhite@srk12.org
Purchase Order # (if applicable):
Quote #:

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
Data Deliverables: [] Level II [] Level III [] Level IV [] EQUIS [] Other
Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other
Date Results Requested: Standard 10 business day
Field Filtered (if applicable): [] Yes [] No
Analysis:
DW PWSID # or WW Permit # as applicable:
Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start) Date	Time	Res. CL2	Number & Type of Containers
SRC-41	DW	G	9/30/25	0645		1 Plastic
SRC-42						
SRC-43				0644		
SRC-44				0641		
SRC-45				0642		
SRC-46				0639		
SRC-47						
SRC-48				0629		
SRC-49				0630		
SRC-50				0627		

Customer Remarks / Special Conditions / Possible Hazards:
Lead
Relinquished by/Company: (Signature) Date/Time: 10/2/25 1200
Relinquished by/Company: (Signature) Date/Time: 10/2 12:10
Relinquished by/Company: (Signature) Date/Time: 10/3/25 600
Relinquished by/Company: (Signature) Date/Time:

Additional Instructions from Pace*:
Thermometer ID: 78211 Correction Factor (°C): 14.8 Obs. Temp. (°C): 14.8 Corrected Temp. (°C): 14.9
Tracking Number:
Date/Time: 10/2/25 12:00
Date/Time:
Delivered by: [] In-Person [] Courier [] FedEx [] UPS [] Other
Page: 5 of 1

CHAIN-OF-CUSTODY Analytical Request Document
Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields



Scan QR code for instructions

Company Name:	FEH BOCES Salmon River CSD	Contact/Report To:	Glen Johnson
Street Address:	637 Co Rd 1, Fort Covington, NY 12937	Phone #:	5672429783
		E-Mail:	gjohnson@srk12.org
		Cc E-Mail:	
Customer Project #:		Invoice To:	Ashley White
Project Name:	Salmon River Central School District	Invoice E-Mail:	acwhite@srk12.org
Site Collection Info/Facility ID (as applicable):		Purchase Order # (if applicable):	
		Quote #:	

Time Zone Collected:	[] AK [] PT [] MT [] CT [X] ET	Country / State origin of sample(s):	New York
Data Deliverables:		Regulatory Program (DW, RCMA, etc.) as applicable:	NY Lead in School DW
[] Level II [] Level III [] Level IV		DW PWSID # or WW Permit # as applicable:	
[] EQUIS		Field Filtered (if applicable):	[] Yes [] No
[] Other		Analysis:	

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Composite End	Res. CL2	Number & Type of Containers		200.8 Drinking Water (Pb only)	Sample Comment
			Date	Time			Plastic	Glass		
SRC-51	DW	G	9/30/25	0628			1		X	
SRC-52				0626						
SRC-53				0625						
SRC-54				0624						
SRC-55				0623						
SRC-56				0622						
SRC-57				0619						
SRC-58				0620						
SRC-59				0621						
SRC-62				0614						

Customer Remarks / Special Conditions / Possible Hazards:	Collected by: Lauren Doherty	Additional Instructions from Pace®:	
Lead	Printed Name: Lauren Doherty	Thermometer ID: 74211	Correction Factor (°C): +0.1
	Signature: [Signature]	Obs. Temp. (°C): 14.8	Corrected Temp. (°C): 14.9
Relinquished by/Company: (Signature)	Date/Time: 10/2/25 12:00	Tracking Number:	
Relinquished by/Company: (Signature)	Date/Time: 10/2/25 12:10		
Relinquished by/Company: (Signature)	Date/Time: 10/3/25 6:00		
Relinquished by/Company: (Signature)	Date/Time:		

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/LogIn Label Here



Scan QR Code for instructions

Company Name: FEH BOCES Salmon River CSD
Street Address: 637 Co Rd 1, Fort Covington, NY 12937

Contact/Report To: Glen Johnson
Phone #: 5672429783
E-Mail: gjohnson@srk12.org
Cc E-Mail:

Customer Project #:

Project Name: Salmon River Central School District

Site Collection Info/Facility ID (as applicable):

Invoice To: Ashley White
Invoice E-Mail: awhite@srk12.org

Purchase Order # (if applicable):

Quote #:

Time Zone Collected: [] AK [] MT [] PT [] CT [] ET

Data Deliverables:

Level	Level II	Level III	Level IV
[] EQUIS			
[] Other			

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other

Date Results Requested: Standard 10 business day

Field Filtered (if applicable): [] Yes [] No

Analysis:

*** Matrix Codes (Insert in Matrix box below):** Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start) Date	Time	Res. CL2	Number & Type of Containers	Plastic	Glass
SRC-83	DW	G	10/21/25	0608		1		
SRC-84				0640				
SRC-85				0640				
SRC-86				0641				
SRC-87				0641				
SRC-88				0625				
SRC-89				0625				
SRC-90				0627				
SRC-91				0621				
SRC-94				0616				

Additional Instructions from Pace®:

Collected By: Lauren Doherty
Printed Name:
Signature:

Received by/Company: (Signature)
Received by/Company: (Signature)
Received by/Company: (Signature)
Received by/Company: (Signature)

Additional Instructions from Pace®:

Coolers: Thermometer ID: 19211 Correction Factor (°C): 19.8 Obs. Temp. (°C): 19.8 Corrected Temp. (°C): 19.9

Tracking Number: 192125 1200

Date/Time: 10/21/25 12:00
Date/Time: 10/21/25 12:10
Date/Time: 10/21/25 12:10
Date/Time: 10/21/25 12:10

Delivered by: [] In-Person [] Courier
Date/Time: 10/21/25 600
Date/Time: 10/21/25 600
Date/Time: 10/21/25 600
Date/Time: 10/21/25 600

Scan QR Code for instructions

Page: 9 of 11



Company Name:	FEH BOCES Salmon River CSD	Contact/Report To:	Glen Johnson
Street Address:	637 Co Rd 1, Fort Covington, NY 12937	Phone #:	5672429783
		E-Mail:	gjohnson@srk12.org
		Cc E-Mail:	
Customer Project #:		Invoice To:	Ashley White
Project Name:	Salmon River Central School District	Invoice E-Mail:	acwhite@srk12.org
Site Collection Info/Facility ID (as applicable):		Purchase Order # (if applicable):	
		Quote #:	

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET	County / State origin of sample(s):	New York
Data Deliverables:		
[] Level II [] Level III [] Level IV	Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW	
[] EQUIS	Rush (pre-approval required): [] 2 Day [] 3 day [] 5 day [] Other _____	
[] Other _____	Date Results Requested: Standard 10 business day	DW PWSID # or WW Permit # as applicable: Field Filtered (if applicable): [] Yes [] No Analysis:

Customer Sample ID	Matrix *	Comp / Grab	Collected		Composite End		Res. CL2	Number & Type of Containers	
			or Composite Start		Date	Time		Plastic	Glass
			Date	Time					
SRC-95	DW	G	10/2/25	0614				1	
SRC-96				0614					
SRC-99				0631					
SRC-100				0631					
SRC-103				0646					
SRC-104				0645					
SRC-105				0649					
SRC-110				0559					
SRC-111				0606					
SRC-112				0604					

Customer Remarks / Special Conditions / Possible Hazards:	Lead	Relinquished by/Company: (Signature)	Date/Time:	Collected By:	Printed Name:	Signature:	Received by/Company: (Signature)
		<i>[Signature]</i>	10/21/25 1200	<i>[Signature]</i>	Lauren Doherty	<i>[Signature]</i>	<i>[Signature]</i>
		<i>[Signature]</i>	10/22 12:10	<i>[Signature]</i>			<i>[Signature]</i>
		<i>[Signature]</i>	10/23/25 1200	<i>[Signature]</i>			<i>[Signature]</i>
		<i>[Signature]</i>		<i>[Signature]</i>			<i>[Signature]</i>

Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY- Affix Workorder/Login Label Here



Company Name:	FEH BOCES Salmon River CSD	Contact/Report To:	Glen Johnson
Street Address:	637 Co Rd 1, Fort Covington, NY 12537	Phone #:	5672429783
		E-Mail:	gjohnson@srk12.org
		Cc E-Mail:	
Customer Project #:		Invoice To:	Ashley White
Project Name:	Salmon River Central School District	Invoice E-Mail:	acwhite@srk12.org
Site Collection Info/Facility ID (as applicable):		Purchase Order # (if applicable):	
		Quote #:	

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET	County / State origin of sample(s):	New York
Data Deliverables:		
[] Level II [] Level III [] Level IV	Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW	
[] EQUIS	Rush (Pre-approval required): [] 12 Day [] 3 day [] 5 day [] Other _____	
[] Other _____	Date Results Requested: Standard 10 business day	
Field Filtered (if applicable): [] Yes [] No Analysis:		
* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Biossary (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SD), Sludge (SL), Caulk		

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers	
			Date	Time	Date	Time		Plastic	Glass
SRC-113	DW	G	10/12/25	0610				1	
SRC-114				0610					
SRC-115				0610					
SRC-116				0624					
SRC-117				0624					
SRC-118				0625					
SRC-119				0625					
SRC-120				0628					
SRC-121				0633					
SRC-122				0646					

Customer Remarks / Special Conditions / Possible Hazards:		Lead	
Relinquished by/Company: (Signature)		Date/Time: 10/2/25 1200	Collected By: Lauren Doherty Printed Name: Signature: 
Relinquished by/Company: (Signature)		Date/Time: 10/2 12:00	Received by/Company: (Signature) 
Relinquished by/Company: (Signature)	115103256	Date/Time:	Received by/Company: (Signature) AEB
Relinquished by/Company: (Signature)		Date/Time:	Received by/Company: (Signature) PACE VI

[illegible]

Effective Date: 4/12/2024

Client Name:

Salmon River

Project #

WO#: 70383633

PM: MC1

Due Date: 10/17/25

CLIENT: SALMON RIVER

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☒ Pace ☐ Other

Tracking #:

Custody Seal on Cooler/Box Present: ☐ Yes ☒ No Seals intact: ☐ Yes ☒ No Temperature Blank Present: ☐ Yes ☒ No
Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☐ Ziploc ☐ None ☒ Other Type of Ice: Wet Blue NoneThermometer Used: TH11 Correction Factor: to 1 14.9 ☒ Samples on ice, cooling process has begun
Cooler Temperature(°C): 14.8 Cooler Temperature Corrected(°C): 14.9 Date/Time 5035A kits placed in freezer

Temp should be above freezing to 5.0°C

USDA Regulated Soil ☒ N/A, water sample)Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? ☐ Yes ☒ NoDid samples originate from a foreign source including Hawaii and Puerto Rico)? ☐ Yes ☒ No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents:

ASB 10/3/23

	COMMENTS:
Chain of Custody Present: ☒ Yes ☐ No	1.
Chain of Custody Filled Out: ☒ Yes ☐ No	2.
Chain of Custody Relinquished: ☒ Yes ☐ No	3.
Sampler Name & Signature on COC: ☒ Yes ☐ No ☐ N/A	4.
Samples Arrived within Hold Time: ☒ Yes ☐ No	5.
Short Hold Time Analysis (<72hr): ☒ Yes ☐ No	6.
Rush Turn Around Time Requested: ☐ Yes ☒ No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) ☒ Yes ☐ No	8.
Correct Containers Used: ☒ Yes ☐ No	9.
-Pace Containers Used: ☒ Yes ☐ No	10.
Containers Intact: ☒ Yes ☐ No	11.
Filtered volume received for Dissolved tests ☐ Yes ☒ No ☐ N/A	Note: if sediment is visible in the dissolved container,
Sample Labels match COC: ☒ Yes ☐ No	12.
-Includes date/time/ID/Analysis Matrix: SL <u>WT</u> OIL OTHER	

Date and Initials of person checking preservation:

ASB 10/3/23

All containers needing preservation have been ☒ Yes ☐ No ☐ N/A	13. ☐ HNO ₃ ☐ H ₂ SO ₄ ☐ NaOH ☐ HCl
pH paper Lot # <u>20528</u>	Sample #
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH>9 Sulfide ☒ Yes ☐ No ☐ N/A NAOH>12 Cyanide)	Initial when completed: Lot # of added preservative: Date/Time preservative added:
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis	14.
Samples checked for dechlorination: ☐ Yes ☐ No ☒ N/A	Positive for Res. Chlorine? Y N
KI starch test strips Lot #	15.
Residual chlorine strips Lot #	Positive for Sulfide? Y N
SM 4500 CN samples checked for sulf ☐ Yes ☐ No ☒ N/A	16.
Lead Acetate Strips Lot #	17.
Headspace in ALK Bottle (>6mm): ☐ Yes ☐ No ☒ N/A	
Headspace in VOA Vials (>6mm): ☐ Yes ☐ No ☒ N/A	
Trip Blank Present: ☐ Yes ☐ No ☒ N/A	
Trip Blank Custody Seals Present ☐ Yes ☐ No ☒ N/A	

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted:

Date/Time:

Comments/ Resolution:

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.